

■ Point of Care

Dear I-STAT user:

The i-STAT® Portable Clinical Analyzer (when powered by alkaline batteries), Electronic Simulator, and cartridges pose no health or safety threat to the operator and, therefore, require neither OSHA Hazard Communications Warning labeling nor Material Safety Data Sheets.

The analyzer can be powered by two 9 volt lithium batteries which can be purchased from i-STAT. The manufacturer of the lithium batteries (Ultralife) provides an MSDS, a copy of which is included. The Electronic Simulator is powered by the analyzer.

The cartridges contain no hazardous materials as can be ascertained by reading the product labeling. The cartridges, once filled with patient blood samples, should be handled with precautions to protect the operator from blood borne pathogens.

If you need further information, please call i-STAT's Technical Support Department at (800) 366-8020, option 1.

Sincerely,

Carol Mc Graw, BSMT (ASCP) Technical Support Specialist

Abbett POC Division 104 Windsor Center Drive East Windsor, NJ 08520

East Windsor, NJ | 609-469-0361

609-469-0260 (fax) carol.mcgraw(01-stat.com

-STAT MediSense

104 Windson Contor Drive Base Windson, NJ 08529 Tel: 005-443-9300

From 609-443-2310

Web; http://www.i-STAT.com